

CERTIFICATE OF CONFORMITY (CERTIFICAT DE CONFORMITÉ)

müller+krempel 

14.210.02

	Customer (Client) ELPACK
--	-----------------------------

Version (Version)	MATIERE CAP / TOP / BOTTOM (Raw Material) (*)	PLUG	VALVE	CAP	PACKAGING	IRRADIATION
	HDPE SABIC PCG 453 00900 (MELIFLEX)	Pure Flow 10	Valve diam. 2,4	CAP white	Double bags	Non Irradiated
BATCH N° (N° de lot)			10975638			
Quantity supplied (Quantité livrée)			48000			
Samples for customer (Echantillons client) Sampling completed according to ISO standard 2859-1, simple plan, normal inspection, level I. (Echantillonnage réalisé selon la norme ISO 2859-1, plan d'échantillonnage simple, contrôle normal, niveau I)			200			
Assembly period (Dates de production)			09/11/2016			

(*) Raw materials are described in the current version of Nemera Specification.

Semi-finished parts (Produits semi-finis)

Article : Composant)	Batch N° (N° de lot)	Bag to Bag (Sac à sac)
CAP	10973101	1 TO 40
TOP	10972991	1 TO 40
BOTTOM	10973588	1 TO 40
VALVE	S130370127	1 TO 40
SPRING	S170220131	1 TO 40
PLUG	S160050004 / S160050005	1 TO 32 / 32 TO 40

INSPECTION (Inspection)	RESULTS (Résultats)
Injected components (Dimensional / Functional / Aspect) Composants injectés (Dimensionnel / Fonctionnel / Aspect)	ACCEPTED (ACCEPTE)
Dimensional (Dimensionnel)	ACCEPTED (ACCEPTE)
Functional (Fonctionnel)	ACCEPTED (ACCEPTE)
Aspect (Aspect)	ACCEPTED (ACCEPTE)
Certificate of irradiation (Certificat d'irradiation)	NOT REQUIRED (NON REQUIS)

Comments (Commentaires) :

This batch was manufactured and inspected according to current version of Nemera specification
(Ce lot a été fabriqué et testé suivant les spécifications Nemera définies dans le cahier des charges en vigueur)

Decision (Décision)	ACCEPTED (ACCEPTE)
Date (Date)	22/11/2016
Name (Nom)	BOISSIERES Anne
Function (Fonction)	Support Qualité Production
Signature Quality Insurance (Vise Assurance Qualité)	